

## ENTREPRENEURSHIP DEVELOPMENT CENTER

Amendment to Tender No.: **EDC/NIT/2020/002**

Item Description – Procurement of Fourier Transformation Infrared Spectrometer

A press tender notice was issued on 4<sup>th</sup> May 2020 for the procurement of Fourier Transformation Infrared Spectrometer under two-bid system. The following amendments were recommended by the Technical Committee.

Specification Ref.	Existing Nomenclature	Amended Nomenclature
Chapter 1 Page 5 Point (v)	Demand Draft/ Bank Guarantee for Rs. 1,25,000 (Rs. One Lakh Twenty-Five Thousand only) / US \$ 1700 (US Dollars Seventeen Hundred Only) towards Earnest Money Deposit. Demand Draft/ Bank Guarantee shall be obtained from Nationalized/ Scheduled Bank only.	Demand Draft/ Bank Guarantee for Rs. 1,25,000 (Rs. One Lakh Twenty-Five Thousand only) / US \$ 1700 (US Dollars Seventeen Hundred Only) towards Earnest Money Deposit. Demand Draft/ Bank Guarantee shall be obtained from Nationalized/ Scheduled Bank only. <b>Due to the existing lock down imposed by the Government from 24<sup>th</sup> March to 17<sup>th</sup> May 2020 due to COVID-19, we will be accepting the EMD via Electronic transfer also to the following Bank Account :</b> <b>Entrepreneurship Development Center</b> <b>Axis Bank, Baner Branch, Pune - 411008.</b> <b>Savings A/C No.: 338010100042936</b> <b>RTGS/NEFT/IFSC: UTIB0000338</b>
<b>Chapter 4</b> <b>Sr. No. 4</b>	The Interferometer must be a dynamically aligned, non-air bearing/air bearing/any other advanced type Michelson design. Non-air bearing one will be given priority	The Interferometer can be dynamically aligned or permanently aligned.

<p><b>Chapter 4</b> <b>Sr. No. 15</b></p>	<p>1) CaF2 cell 6 – 8 micron – 2 nos., temperature controller (5 to 95 deg C) suitable for CaF2 2) All other accessories essential to carry out the applications stated above</p>	<p>1) CaF2 cell 6 – 8 micron – 2 nos., temperature controller (5 to 95 deg C) suitable for CaF2 2) ATR accessory with diamond crystal, with provision for temperature control i.e. from 15<sup>o</sup>C to 110<sup>o</sup>C. The ATR should preferably be manufactured by the Principal Company. If an ATR from a third party is quoted, please provide all details including the required accessories &amp; respective software. 3) All other accessories essential to carry out the applications stated above.</p>
---	---	--

The terms and conditions of the notice issued on our EDC website [www.venturecenter.co.in](http://www.venturecenter.co.in) will remain unchanged.

15.05.2020

General Manager (EDC)