

ENTREPRENEURSHIP DEVELOPMENT CENTER

Amendment to Tender No.: **EDC/NIT/2020/001**

Item Description – Procurement of Ultra High Performance Liquid Chromatograph

A press tender notice was issued on 4th May 2020 for the procurement of Ultra High Performance Liquid Chromatograph under two-bid system. The following amendments were recommended by the Technical Committee.

Specification Ref.	Existing Nomenclature	Amended Nomenclature
Chapter 4 Sr. No. 5	Essential Parameters: Fluorescence Detector Wavelength range : 200 – 900 nm	Essential Parameters: Fluorescence Detector Wavelength range from 200nm to 750 nm is essential. Wavelength range from 200 nm to 900 nm will be an additional capability.

The terms and conditions of the notice issued on our EDC website www.venturecenter.co.in will remain unchanged.

18.05.2020

General Manager (EDC)