

ENTREPRENEURSHIP DEVELOPMENT CENTER

Clarification on Tender Number- **EDC/NIT/2019/004**

Item Description – Procurement of Multi angle light scattering detector

A press tender notice was issued on 8th November 2019 for the procurement of Multi angle light scattering detector under two-bid system.

A pre-bid meeting was held on 19th November 2019 at 10:00 am and the minutes of the meeting is as under:

- 1) Prebid meeting was attended by representative from Spinco Biotech Pvt. Ltd - distributor for Wyatt Technologies.
- 2) Prior and during the meeting written queries were received from representatives .Our replies to the same are as summarized below:

Specification Ref.	Query	Our reply:
3.10.1	<p>Warranty/Support</p> <p>The items covered by the schedule of requirement shall carry minimum <u>two</u> years of comprehensive warranty from the date of acceptance of the equipment by EDC. Warranty shall include free maintenance of the whole equipment supplied including free replacement of parts. The comprehensive warranty includes onsite warranty with parts. Defects should be attended to on immediate basis and all efforts should be taken to bring the equipment back into operation at the earliest available time.</p> <p>Comprehensive Maintenance Charges for 3rd, 4th and 5th years have to be quoted separately in the bids</p>	<p>Will be amended.</p> <p>Please refer to point 3 in the Minutes below.</p> <p>Visit our website www.venturecenter.co.in to see the amendment document.</p>
3.6	<p>Performance Bank Guarantee</p> <p>The 10% Security Deposit which is mentioned above may be extended as Performance Bank Guarantee for the warranty period of two years from the Date of Installation</p>	<p>Performance Bank Guarantee will as per Clause 3.6 which states that "The 10% Security Deposit which is mentioned above may be extended as Performance Bank Guarantee for</p>

		the warranty period."
3.11.2	Training One week advanced training for two users nominated by EDC on operation, method development, maintenance, software, data interpretation(qualitative and quantitative, unknown identification), etc. at the manufacturer's international application laboratory	Will be amended. Please refer to point 4 in the Minutes below. Visit our website www.venturecenter.co.in to see the amendment document.
Aim	A Multi angle light scattering detector for biopharma characterization is required. The system shall be upgradeable in future with a Dynamic Light Scattering module for carrying out simultaneous SEC-MALS-DLS studies using the same light source and flow cell.	Will be amended. Please refer to point 5 in Minutes below. Visit our website www.venturecenter.co.in to see the amendment document.
Sr.No: 3	System should have laser light source with controllable power, support laser wavelengths of 658 nm and high -gain, high range dynamic photodiodes. The system shall have rear laser monitor for stable laser output and also a forward transmission monitor to correct signals for absorbing samples	No amendment
Warranty	2-year comprehensive 'spares-included' warranty. Vendors must mention exhaustively all the parts that are covered under Warranty.	Will be amended. Please refer to point 3 in the Minutes below. Visit our website www.venturecenter.co.in to see the amendment document.
Training Specifications	Training at manufacturer's assembly site for specific hardware training requirements, to be quoted separately in the bids	Will be amended. Please refer to point 4 in the Minutes below. Visit our website www.venturecenter.co.in to see the amendment

		document.
3.11.3	We request that this be deleted	No amendment

3) Clause 3.10.1 is amended as follows:

Existing Nomenclature: The items covered by the schedule shall carry minimum five years of comprehensive warranty from the date of acceptance of the equipment by EDC.

Amended Nomenclature: The items covered by the schedule shall carry two years of comprehensive warranty from the date of acceptance of the equipment by EDC and Annual Maintenance Contract in years 3, 4 & 5.

4) Clause 3.11.2 is amended as follows;

Existing Nomenclature: One week advanced training for two users nominated by EDC on operation, method development, maintenance, software, data interpretation(qualitative and quantitative, unknown identification), etc. at the manufacturer's international application laboratory

Amended Nomenclature: One week advanced training for two users nominated by EDC on operation, method development, maintenance, software, data interpretation(qualitative and quantitative, unknown identification), etc. at the manufacturer's application laboratory in India.

5) Aim in Chapter 4

Existing Nomenclature: A Multi angle light scattering detector for biopharma characterization is required.

Amended Nomenclature: A Multi angle light scattering detector for biopharma characterization is required. The system shall be upgradeable in future with a Dynamic Light Scattering module for carrying out simultaneous SEC-MALS-DLS studies using the same light source and flow cell.

6) In general the bidders expressed their satisfaction about the details and Specifications given in the bid document.

The terms and conditions of the notice issued on our EDC website www.venturecenter.co.in will remain unchanged.

21.11.19

General Manager (EDC)